



QUICKSHIELD Complete Safety Tube Holder and VACUETTE[®] Multiple Use Drawing Needle combination

Instructions for use

Intended Use: The QUICKSHIELD Complete is intended to be used only with VACUETTE[®] Blood Collection Tubes as a system in routine venipuncture procedures. This device is designed with a safety shield, which can be activated to cover the needle immediately following venipuncture to aid in the protection against accidental needlestick injury. This device is to be used by properly trained healthcare professionals only in accordance with the instructions.

Product Description: The sterile QUICKSHIELD Complete is a single packed QUICKSHIELD Safety Tube Holder with pre-threaded VACUETTE[®] Multiple Use Drawing Needle in a blister pack. The QUICKSHIELD Complete is designed with a safety shield, which can be activated to cover the needle immediately following venipuncture to aid in the protection against accidental needlestick injury.

Precautions/Cautions

Precautions

Check QUICKSHIELD Complete before use. Do not use QUICKSHIELD Complete when damaged! Any QUICKSHIELD Complete that becomes contaminated with blood is considered hazardous and must be disposed of immediately. After venipuncture, dispose of QUICKSHIELD Complete as one unit in appropriate sharps container in accordance with policies and procedures of your facility.

Keep hands behind needle at all times during use and disposal.

Caution

1. Handle all biological samples and blood collection “sharps” (lancets, needles, Luer adapters, and blood collection sets) according to the policies and procedures of your facility.
2. Obtain appropriate medical attention in the case of any exposure to biological samples (for example, through a puncture injury), since they may transmit HIV (AIDS), viral hepatitis, or other infectious disease.
3. Discard all blood collection “sharps” in biohazard containers approved for their disposal.
4. Gloves should be worn at all times during venipuncture to minimize exposure hazard.

Storage

Store holders at 4–25°C (40–77° F).

NOTE: Avoid exposure to direct sunlight. Exceeding the maximum recommended storage temperature may lead to impairment of the holder.

Handing Procedures

READ THIS ENTIRE CIRCULAR BEFORE PERFORMING VENIPUNCTURE.

WEAR GLOVES DURING VENIPUNCTURE AND WHEN HANDLING BLOOD COLLECTION TUBES TO MINIMISE EXPOSURE HAZARD!

1. Open the blister packaging and fully remove device from packaging.

NOTE: The blister packaging is an indicator of sterility and product integrity. If the blister packaging is broken or damaged, dispose of the QUICKSHIELD Complete and select another.

NOTE: Do not bend needles or use QUICKSHIELD Complete when needle is bent.

3. Select the puncture site. Apply the tourniquet (max. 1 minute). Prepare the venipuncture site with appropriate antiseptic. **DO NOT PALPATE THE VENIPUNCTURE AREA AFTER CLEANSING!**

4. Place the patient's arm in a downward position. Remove the needle cap. Perform the venipuncture, with the patient's arm downward and the tube cap upper-most.

5. Push the tube into the QUICKSHIELD Complete and onto the needle valve puncturing the rubber diaphragm. Centre tubes in the QUICKSHIELD Complete when penetrating the cap to prevent sidewall penetration and subsequent premature vacuum loss.

6. REMOVE THE TOURNIQUET AS SOON AS BLOOD APPEARS IN THE TUBE. DO NOT ALLOW THE CONTENTS OF THE TUBE TO CONTACT THE CAP OR THE END OF THE NEEDLE DURING THE PROCEDURE. i.e. under no circumstances should the tube be turned upside down during the procedure. Always hold in place by pressing the tube with the thumb to ensure complete vacuum draw.

NOTE: Blood may occasionally leak from the needle sleeve. Practice universal safety precautions to minimise hazard exposure.

7. Place succeeding tubes in the holder.

8. As soon as blood stops flowing into the last tube, carefully remove the needle from the vein, applying pressure to the puncture site with a dry sterile swab until bleeding stops. Once clotting has occurred, apply an adhesive plaster or bandage if desired.





NOTE: After venipuncture the top of the cap may contain residual blood. Take proper precautions when handling tubes to avoid contact with this blood.

Activation of safety shield

Immediately after removing needle from the vein the following procedures are used to activate the safety shield to prevent needlestick injury.

1. Activate the safety shield by gently pressing the shield towards the needle on a stable surface. (See A)

2. An audible click is made ensuring the user the safety shield has been properly and fully activated. The risk of a needlestick injury is thus virtually eliminated. (See B)



3. Dispose of the entire QUICKSHIELD Complete device in an appropriate disposal container. (See C)

Disposal

The general hygiene guidelines and legal regulations for the proper disposal of infectious material should be considered and followed.

Disposable gloves may prevent the risk of infection.

Contaminated holders must be disposed of immediately after blood collection in suitable biohazard disposal containers, which can then be autoclaved and incinerated.

Label Information on the inner packaging (bags) and the outer carton.

Ref. Reference Number: QUICKSHIELD Complete can be ordered using this item numbers

LOT Lot Number: batch number



Single use product



Expiry Date: shows expiration date of QUICKSHIELD Complete.



greiner bio-one